

भारतसरकार
क्षेत्रीय औषध परीक्षणप्रयोगशाला

केन्द्रीय औषध मानकनियन्त्रण संगठन
(स्वास्थ्य सेवाओंका महानिदेशालय)
स्वास्थ्य एवंपरिवारकल्याणमंत्रालय
सैक्टर-39 सी, चन्डीगढ़-160036
फोननं: 0172-2688239, फैक्सनं: 0172-2636316



NABL

GOVERNMENT OF INDIA
REGIONAL DRUGS TESTING LABORATORY

CENTRAL DRUGS STANDARD CONTROL ORGANISATION
(Directorate General of Health services)

Ministry of Health & Family Welfare

SECTOR-39C, CHANDIGARH – 160036

Tel. No: -0172-2688239, Fax- No: 0172-2636316

E-mail id: rdtlchd@cdsco.nic.in

ACCREDITED LABORATORY

No. 3-15 /2020-21/St/Equipments/Part-I/2294 /20

Dated: 24.12.2020

TENDER NOTICE INVITING QUOTATION

The Director, Regional Drugs Testing Laboratory, Chandigarh invites quotation from the interested and bonafide manufacturer, dealer, distributor or retailer for supplying of one vacuum oven along with pump as per the specifications mentioned in Annexure-A at Regional Drugs Testing Laboratory, Chandigarh.

For detailed notice, please visit <http://www.cdsco.gov.in>

The detailed terms & conditions for supplying of one vacuum oven along with pump are available on the above website.

1. The quotations should be submitted in sealed envelope containing:
 - (i) Acceptance of terms & conditions as per Annexure-I
 - (ii) The Performa at **Annexure-II** and **Annexure-III**, duly filled in along with relevant documents/information's.
2. The sealed envelope should be super scribed with "Quotations for supply of one vacuum oven along with pump". Quotations should be addressed to the Director, Regional Drugs Testing Laboratory, Sector-39C, Chandigarh and be reached at this office within 21 days from the publishing of the notice i.e. 13.01.2021 up to 3:00 PM.
3. The quotations, so received will be opened on same day i.e. on 13.01.2021 at 3:30 PM in the presence of interested quotationers or their authorized representatives.
4. Tender received after stipulated date & time will be rejected forthwith.
5. The competent authority in the RDTL reserves the right to amend any of the terms and conditions in the tender document or reject any or all applications (offers) or cancel the tender without giving any notice or assigning any reason thereof. The decision of competent authority in this regard will be final and binding.

Yours faithfully

Director
RDTL, Chandigarh

Terms and Conditions

1. Rates quoted for the supply and installation of vacuum oven along with pump should be inclusive of supply/freight/taxes etc. and no extra payment will be made thereafter.
2. Only manufacturing firms/authorized distributors are eligible to participate. While Submitting the Tenders, the tendering firm shall have to furnish the proof of PAN No., Service Tax No., Sales Tax No., etc. as per Annexure-II.
3. The quotations duly filled in as per the instruction giving herein may be submitted by hand or by post to this office up to 3:00 PM on dated 13.01.2021.
4. The quotation received after due date and time or incomplete in any respect shall be summarily rejected.
5. **The quotations so received will be opened on same day i.e. 13.01.2021 at 3:30 PM in presence of interested quotationers or their authorized representative.**
6. No Advance payment would be made. Payment will be made only after complete supply and satisfactory working of the system.
7. The tenderer should supply and installation of vacuum oven along with pump would be made within 30 days from the date of issue of purchase order.
8. Over-writing on the quotations must be avoided. Otherwise the quotations are likely to be rejected. The RDTL reserves the right to accept or reject any tender, in whole or in part thereof without assigning/ specifying any reason (s) thereof and the decision of the competent authority in the RDTL shall be final and binding on the contracting firm.
9. All the participating firms may ensure that the vacuum oven along with pump are of only Genuine quality and in compliance of Technical specification, if at any stage it is found that the item supplied is of sub-standard quality/inferior quality, the Performance Security will be forfeited and if the cost of the item is not covered from the Performance Security the same will be recovered from the supplier.
10. If in any circumstances, the successful lowest quotationer refuses or unable to supply the item with in stipulated time, the second successful lowest quotationer will be offered.
11. Any dispute arising in the matter shall be resolved through an arbitrator to be nominated by the competent authority in RDTL.
12. The contract shall be subject to Indian Laws and jurisdiction of the court located in Chandigarh.
13. The competent authority of RDTL reserve right to cancel the bid at any stage without assignee any reason.



Director

RDTL, Chandigarh

PROFORMA (Technical bid)

S.No.	Particulars	To be filled by the Tenderer
1.	Name of the agency	
2.	Whether brief profile of the agency is enclosed	
3.	Detailed Office Address of the agency with office telephone No. e-mail address, Fax No., Mobile No. and name/designation of the contract person.	
4.	Whether registered with all concerned Govt. authorities (Register of companies, Commissioner Employees etc.)(Copies of all certificates of Registration should be attached)	
5.	PAN Number (copies to be enclosed)	
6.	TIN Number Registration No. (Copy to be enclosed)	
7.	Service Tax Registration No. (copy to be enclosed)	
10.	Whether copies of income Tax return for last 2 years attached	
11.	Whether a copy of terms and conditions- Annexure-I duly signed as token of acceptance of the' same is attached	
12.	Whether the firm is blacklisted by any Government Department or any criminal case is registered against the firm or its owner/partners anywhere in India (if no, an undertaking to this effect is attached in this regard.)	

(Signature of the authorized person)

Date:

Place:

Name:

Designation:

Seal:

Annexure-III

Financial Bid for Supply and Installation of Vacuum Oven along with pump in RDTL Sector 39-C,
Chandigarh.

With reference to RDTL letter No.....dated.....Regarding tender for supply and
installation of vacuum oven along with pump in RDTL, I/we quote the rate, inclusive of supply/
installation/freight etc. As under:-

S.No	Item Name/Qty	Specifications	Cost of Apparatus
1.	Vacuum Oven along with pump in RDTL		

IMPORTANT:

In case of discrepancy between the unit price and total price, the unit price shall prevail for computation.

Date:

Place:

(Signature of the authorized person)


Name:

Designation:

Seal:

TECHNICAL SPECIFICATION FOR THE VACUUM OVEN ALONG WITH PUMP		
SR.NO.	PARAMETERS	REQUIRMENTS
1.	Capacity	40-50 Ltr. or better
2.	Material	
(i)	Glass door	• Safety Double Glass against implosions
(ii)	Exterior Material	• External Metallic frame with epoxy resin coating in a chamber
(iii)	Interior Material	• In a chamber with inner side door made up of standard Steel 304 or better • Double insulated walls and door with low conductivity glasswool.
3.	Temperature	
(i)	Control Type	Temperature and time control by digital PID microprocessor, with minimum 10 editable programs of customizable name, programmable with up to 10 temperature/time segment
(ii)	Temperature Homogeneity	$\pm 5^{\circ}\text{C}$ or better
(iii)	Temperature range	Minimum 10°C above ambient to $+ 200^{\circ}\text{C}$ or better
(iv)	Temperature Stability	$\pm 1^{\circ}\text{C}$ or better
(v)	Timer	Start and timer stock adjustable by date, hour and minutes
(vi)	Temperature Display	Intuitive alphanumeric touch screen user interface that displays time temperature and cycles in graph, has audio visual alarms and advanced cycle and timer programming.
(vii)	Temperature Sensor	PT100 Class A temperature probe or better
(viii)	Temperature Display accuracy	$\pm 2^{\circ}\text{C}$ or better
4.	Connectivity	Ethernet connection or better
5.	Pressure (Vacuum)	
(i)	Pressure Vacuum	0.01 mbar or better
(ii)	Vacuum display	Dialer gauge or better
(ii)	Pressure Display resolution	0.01 or better
(iv)	Shelf position	2 perforated removable trays made of aluminum with galvanic treatment
6.	Dimensions (W X H X D),in (mm)	
(i)	Chamber Internal	420 x 350 x 348 or better
(ii)	Overall	580 x 570 x 650 or better
7.	Vacuum Pump	
(i)	Caudal flow	70 L/minute or better
(ii)	Pressure(Vacuum)	Vacuum pump with ultimate vacuum 0.002 mbar or better
(iii)	Anti-suck back valve	2 stage vacuum pump with vacuum gauge and non returnable valve to protect the system from oil contamination.
8.	Oil Level indicator	Visor with oil level indicator and oil drain system for easy emptying.
9.	Power	230 volt 50 Hz
10.	Safety with certification	Safety according to EN61010-1, EN61010-2-010

11.	Warranty	Minimum 1 year from date of installation
12.	Validation	Onsite temperature mapping, consisting 8 hours cycle of loading chamber at 105 °C and 0.01 mbar pressure
13.	Documents	Operating Manual IQ/OQ and PQ documents
14.	Training	Onsite training has to be provided



Director